

		<p>±5 dB on three (3) consecutive bands when compared to initial evaluation; and ±2 dB when comparing the average of the absolute differences between initial and continuing qualification evaluation.</p>		<p>Applicable only to Continuing Qualification Evaluations. If frequency response plots are provided for each channel at the initial evaluation, these plots may be repeated at the continuing qualification evaluation with the following tolerances applied:</p> <p>(a) The continuing qualification $\frac{1}{3}$ octave band amplitudes must not exceed ±5 dB for three consecutive bands when compared to initial results.</p> <p>(b) The average of the sum of the absolute differences between initial and continuing qualification results must not exceed 2 dB (refer to table C2C in Appendix C).</p>			X	<p>Measurements are compared to those taken during initial qualification evaluation.</p>
--	--	--	--	--	--	--	---	--